



06/15/11

Recommended Spare Parts

**Model 400 HCLD
Gas Analyzer**

<u>Part Number</u>	<u>Description</u>
110271-2	EPC Valve (Sample)
110773-2	EPC Rebuild Kit (Sample)
110386-1	EPC Valve (Air)
110773-3	EPC Rebuild Kit (Air)
310127-1	Filter Element (Sample)
410291-4	Fuse, 4 Amp 250V (Slo Blo) (230 VAC)
410289-2	Fuse, 8 Amp 250V (Slo Blo) (115 VAC)
310074-1	Long Pass Filter (Red Lens)
110378-1R	NO _x Converter
110083-1	NO _x Converter Carbon Refill Pack
110415-5	NO _x Solenoid Valve Assembly
410307-2	Ozone Lamp
310821-1	Reaction Chamber
310081-1	Reaction Chamber O-Ring
310005-7	Pump, Sample 115V
310005-8	Pump, Sample 230V
310309-8	Rebuild Kit, Pump N012
310309-7	Rebuild Kit, Pump N010
310235-1	Regulator, Air
110487-2	Solenoid Valve Assembly (Span) (Optional)
110487-1	Solenoid Valve Assembly (Zero) (Optional)

SPECIFICATIONS SUBJECT TO CHANGE WITHOUT PRIOR NOTICE
INFO@GASANALYZERS.COM

CALIFORNIA ANALYTICAL INSTRUMENTS, INC., 1312 W GROVE AVE, ORANGE, CA 92865