06/15/11

Recommended Spare Parts

Model 600 NDIR
Gas Analyzer

Part Number	Description
410291-3	Fuse, 2 Amp (Slo Blo) (230 VAC) (requires 2 fuses)
410289-4	Fuse, 3 Amp (¼ x 1¼, 250V, Slo blo) for 115V analyzers
ZZPZFU3-C050	IR Source & Chopper Assembly (Used on analyzers whose serial # does NOT end with the letter M)
110916-1	IR Source & Chopper Assembly (Stepper) (Used on analyzers whose serial # end with the letter M)
110157-16	Pump, Dual Headed
310309-4	Rebuild Kit, Sample Pump
310975-1	Pump, Sample / 115 VAC (Optional) (Used on analyzers whose serial # does NOT end with the letter M)
310975-2	Pump, Sample / 230 VAC (Optional) (Used on analyzers whose serial # does NOT end with the letter M)
110157-15	Pump, Sample / 24 VDC (Optional) (Used on analyzers whose serial # end with the letter M)
310310-1	Sintered Metal Filter Rebuild Kit (O_2 only) (includes Filter Element, Springs, O-Ring) **For Analyzers Ranged:** **CO 0 - 200, 0 - 500, 0 - 1000, 0 - 2000, 0 - 5000 ppm, 1%** **CH_4 1000 - 5000 ppm** **CO_2 0 - 400, 0 – 1000, 0 - 2000 ppm** **SO_2 1000 ppm through 1%**
ZZPZFU3-C310	Pipe Cell Window
ZZPZFU3-C320	Pipe Cell O-Ring **Pipe Cell Liner: The pipe cell liner is the same length as the pipe cell. Specify the corresponding part # from the following for the liner.**
910064	64 mm
910125	125 mm
910250	250 mm **For Analyzers Ranged:** **CO 0 – 1% and Higher** **CH_4 1% and Higher** **CO_2 0 – 2000 ppm and Higher** **SO_2 1% and Higher**
ZZPZFU3-C080	Block Cell Window
ZZPZFU3-C090	Block Cell O-Ring